



# Autoclave Warning



Black base and transparent lids are NOT autoclavable. Only templates are autoclavable.



See instructions for use for more information.



800.552.5512 | [ULTRADENT.COM](http://ULTRADENT.COM)

1008735AR01 040220

© 2020 Ultradent Products, Inc. All Rights Reserved.

